Linearity error:

Oxygen Sensor OOM202

ENVITEC

Measurement Range:0-100 % oxygenOutput in ambient air:13 to 16mV

Electrical Interface: 3pin (Molex 22-11-1031)

Accuracy and Repeatability: < 1 % vol. O2 when calibrated at 100 %

Oxygen < 3 % relative

Response time: < 12 sec. to 90 % of final value **Zero Offset Voltage:** < 200 µV in 100 % nitrogen

applied for 5 min

Cross Interference: < 0.5 % vol. O₂ response to:

10 % CO₂ balance N₂ 80% N₂O balance N₂

7.5% Halothane balance N₂ 7.5 % Isoflurane balance N₂ 7.5 % Enflurane balance N₂ 9% Sevoflurane balance N₂ 20% Desflurane balance N₂

Influence of Humidity: - 0.03 % rel. per % RH at 25°C

Influence of Pressure: proportional to change in oxygen partial

pressure

Influence of Mechanical Shock: < 1% relative after a fall from 1m

Operating Temperature: 0 to 50°C

Temperature Compensation: built-in NTC compensation

Effect of Temperature between +25°C and +40°C: 3 % relative error **Compensation (steady state):** between 0 °C and +50 °C: 8 % relative error

Operating Humidity: 0-99 % RH non-condensing **Long Term Output Drift:** < 1 % vol oxygen per month

typically < - 15 % relative over lifetime


Storage Temperature: $-20 \text{ to } +50 ^{\circ}\text{C}$ Recommended Storage: $+5 \text{ to } +15 ^{\circ}\text{C}$ Recommended Load: $\geq 10 \text{ kOhms}$

Warm-Up Time: < 30 minutes, after replacement of sensor

Nominal Sensor Lifetime: ≥ 1.000 000 % vol oxygen hours approximately 28 grams

Part No.: 01-00-0047

All specifications are applicable at standard conditions: 1013 hPa, 25°C dry ambient air



Use the advantages:

- Meet DIN EN 21647
- Designed and manufactured according to EN ISO 9001 : 2000 and EN ISO 13485 : 2003
- Accurate and reliable response
- Resistant to N₂O
- Excellent signal stability
- High product quality
- Short lead times
- Technical support



ENVITEC- WISMAR GMBH

Alter Holzhafen 18 D-23966 Wismar

Phone: +49-(0) 3841- 360 1 Fax: +49-(0) 3841- 360 222 E-Mail: <u>info@envitec.com</u> <u>http://www.envitec.com</u>